

70257 Radiological Service Equipment and Supplies

(a)

There shall be sufficient equipment and supplies maintained to adequately perform the radiological services that are offered in the hospital. As a minimum, the following equipment shall be available: (1) At least one radiographic and fluoroscopic unit. On and after January 1, 1977, fluoroscopic units shall be equipped with image intensifiers. (2) Film processing equipment.

(1)

At least one radiographic and fluoroscopic unit. On and after January 1, 1977, fluoroscopic units shall be equipped with image intensifiers.

(2)

Film processing equipment.

(b)

Proper resuscitative and monitoring equipment shall be immediately available.