

70621 Respiratory Care Service Equipment and Supplies

(a)

There shall be sufficient types and quantity of equipment to provide the appropriate inhalation of the several gases, aerosols and such other modalities required for the anticipated nature and variety of procedures.

(b)

Equipment shall be calibrated in accordance with manufacturer's instructions and records of such calibrations shall be kept.