

L701

(Rev. 210; Issued:02-03-23; Effective:02-03-23; Implementation:02-03-23)

§418.106(f) Standard: Use and maintenance of equipment and supplies

- (1) The hospice must ensure that manufacturer recommendations for performing routine and preventive maintenance on durable medical equipment are followed. The equipment must be safe and work as intended for use in the patient's environment. Where a manufacturer recommendation for a piece of equipment does not exist, the hospice must ensure that repair and routine maintenance policies are developed. The hospice may use persons under contract to ensure the maintenance and repair of durable medical equipment.**
-

