

**D6055**

*(Rev. 166, Issued: 02-03-17, Effective: 03-03-17, Implementation: 03-03-17)*

**§493.1413 Standard; Technical consultant responsibilities**

**(b)(9) unless test methodology or instrumentation changes, in which case, prior to reporting patient test results, the individual's performance must be reevaluated to include the use of the new test methodology or instrumentation.**