

D5785

(Rev. 140, Issued: 05-29-15, Effective: 05-29-15, Implementation: 05-29-15)

§493.1282 Standard: Corrective actions

(b)(3) The criteria for proper storage of reagents and specimens, as specified under §493.1252(b), are not met.

Probes §493.1282(b)(3)

What action does the laboratory take if the storage temperature for a test system's reagents falls outside the acceptable limits?