

**D5467**

*(Rev. 166, Issued: 02-03-17, Effective: 03-03-17, Implementation: 03-03-17)*

*§493.1256 Standard: Control procedures*

**(d)(9) When using calibration material as a control material, use calibration material from a different lot number than that used to establish a cut-off value or to calibrate the test system.**