

**D2072**

**(Rev. 140, Issued: 05-29-15, Effective: 05-29-15, Implementation: 05-29-15)**

**§493.835 Standard; Syphilis serology**

**(d)(1) For any unsatisfactory testing event for reasons other than a failure to participate, the laboratory must undertake appropriate training and employ the technical assistance necessary to correct problems associated with a proficiency testing failure.**

**(2) For any unacceptable testing event score, remedial action must be taken and documented, and the documentation must be maintained by the laboratory for two years from the date of participation in the proficiency testing event.**