

**D2087**

**§493.841 Standard: Routine Chemistry**

**(a) Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.**