

**D6044**

*(Rev. 166, Issued: 02-03-17, Effective: 03-03-17, Implementation: 03-03-17)*

**§493.1413 Standard; Technical consultant responsibilities**

**(b)(6) Ensuring that patient test results are not reported until all corrective actions have been taken and the test system is functioning properly;**