

## **D5449**

*(Rev. 166, Issued: 02-03-17, Effective: 03-03-17, Implementation: 03-03-17)*

### ***§493.1256 Standard: Control procedures***

**(d)(3)(ii) Each qualitative procedure, include a negative and positive control material;**

***Interpretive* Guidelines §493.1256(d)(3)(ii)**

#### **Urinalysis**

Photomicrographs or charts of all possible urine sediment components will meet the control requirement for manual microscopic urinalysis examinations. Use D5445.