

D2109

(Rev. 140, Issued: 05-29-15, Effective: 05-29-15, Implementation: 05-29-15)

§493.845 Standard; Toxicology

Refer to Subpart I for analytes or tests for which laboratory PT performance is to be evaluated, which include serum, plasma, or blood.

(a) Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.