

D2162

(Rev. 140, Issued: 05-29-15, Effective: 05-29-15, Implementation: 05-29-15)

§493.859 Standard; ABO group and D (Rho) typing

(f) Failure to achieve satisfactory performance for the same analyte in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.