

D5733

(Rev. 166, Issued: 02-03-17, Effective: 03-03-17, Implementation: 03-03-17)

§493.1278 Standard: Histocompatibility

(a)(3) Reagent typing sera inventory prepared in-house must indicate source, bleeding date and identification number, reagent specificity, and volume remaining.