

**D2153**

*(Rev. 166, Issued: 02-03-17, Effective: 03-03-17, Implementation: 03-03-17)*

**§493.859 Standard: ABO Group and D (Rho) Typing**

**(a) Failure to attain a score of at least 100 percent of acceptable responses for each analyte or test in each testing event is unsatisfactory analyte performance for the testing event.**