

**D2163**

**(Rev. 140, Issued: 05-29-15, Effective: 05-29-15, Implementation: 05-29-15)**

**§493.859 Standard; ABO group and D (Rho) typing**

**(g) Failure to achieve an overall testing event score of satisfactory for two**

**consecutive testing events or two out of three consecutive testing events is  
unsuccessful performance.**