

D2127

(Rev. 166, Issued: 02-03-17, Effective: 03-03-17, Implementation: 03-03-17)

§493.851 Standard: Hematology

(d) Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.