

## **70257 Radiological Service Equipment and Supplies**

### **(a)**

There shall be sufficient equipment and supplies maintained to adequately perform the radiological services that are offered in the hospital. As a minimum, the following equipment shall be available: (1) At least one radiographic and fluoroscopic unit. On and after January 1, 1977, fluoroscopic units shall be equipped with image intensifiers. (2) Film processing equipment.

#### **(1)**

At least one radiographic and fluoroscopic unit. On and after January 1, 1977, fluoroscopic units shall be equipped with image intensifiers.

#### **(2)**

Film processing equipment.

### **(b)**

Proper resuscitative and monitoring equipment shall be immediately available.