

## 75200 Facilities

### (a)

All dialyzer reprocessing procedures and tests, except rinsing of the dialyzer, shall be performed in a room separate from the dialysis treatment area, or in the patient treatment area if no patients are present or as otherwise noted in this Article.

### (b)

The reprocessing room or area shall be restricted to authorized personnel.

### (c)

For the purpose of minimizing the potential for dialyzer/patient mixup and cross-contamination, the reprocessing area shall be of sufficient size to allow for dialyzer reprocessing, storage, and testing.

### (d)

The reprocessing area shall be maintained in a clean and sanitary manner, and shall have a method of disposal of air and chemical effluents which will not contaminate patients.

### (e)

Air levels of disinfectants in the reprocessing area and in the patient treatment area shall be in compliance with California Administrative Code, Title 8, Section 5155, "Airborne Contaminants."

### (f)

Reprocessing rooms and/or the reprocessing area shall be in compliance with the ventilation requirements of odorous and patient treatment rooms specified in the California Administrative Code, Title 24, Chapter 4-21, Part 4, "Air Conditioning, Heating, and Ventilation Systems for Hospitals."